



TransCanada Calibrations Ltd.
 Box 880
 Ile des Chenes, Manitoba, Canada
 ROA 0T0
 (204) 878-4373



ISO/IEC 17025 Accredited
 CLAS Certification No. 2005-04
 SCC Accredited Lab No.: 591

Certificate of Calibration

Customer:	Lakeside Process Controls 5250 Orbitor Drive Mississauga, ON L4W 5G7 Enbridge Gas	Certificate Number:	3051
		Work Order:	08-101-227-01
		Date of Calibration:	November 17, 2008
		Units	Metric
End User:			
Meter Under Test			
Type:	Ultrasonic	Medium:	Natural Gas
Manufacturer:	Daniel	Pressure: (kPa)	6152.95
Model:	Senior Sonic	Temperature: (°C)	27.35
Serial Number:	08-400217	Density: (kg/m3)	45.35
Meter ID (m)	0.2545	Compressibility:	0.90135
Year of Manufacture:	2008	Direction:	Unidirectional
Meter Details			
CPU Board S/N:	6659	Pulses/m3:	3178.32
CPU Software Ver:	1.61	Meter output tested:	Frequency
Tag Number:	N/A		
Pipe Spool S/N:	N/A		
Run Configuration:	UST 1270mm, CPA 50E, UST 2540mm, METER, DST 1120.9mm		
Flow Conditioning:	CPA 50E Type A		
Flow Conditioner S/N:	17032012		
Witness:	None		
Comments:	None		

Test Method: All calibrations are performed as outlined in TCC calibration procedure GC1-5
 The actual flow of the gas meter is established with the aid of standard gas meters, pressure and temperature transmitters.
 In determination of this flow rate the pressure measurement point noted with Pr, is normatively established.

Results: The measurement results are stated on page 2 of 2. The type A uncertainty in the measurement is based on 2x the standard deviation (2XSTDEV). The total expanded uncertainty (U_{tot}) can be calculated as: $U_{tot} = \text{SQRT}[(U_{facility})^2 + (2XSTDEV)^2]$


Traceability: The Calibration Laboratory Assessment Service (CLAS) of the National Research Council of Canada (NRC) has assessed and certified specific calibration capabilities of this laboratory and traceability to the International System of Units (SI) or to standards acceptable to the CLAS program. The flow meter under test was calibrated in comparison with TransCanada Calibrations flow primary standards which are traceable to the SI and to recognized national metrology institutes. This certificate of calibration is issued in accordance with the conditions of certification granted by CLAS and the conditions of accreditation granted by the Standards Council of Canada (SCC). Neither CLAS nor SCC guarantee the accuracy of individual calibrations by accredited laboratories.

Facility Expanded Uncertainty, k=2:	Flow Rate	$U_{facility}$ [% of reading]
1,000	to 50,000 m3/h	0.20 %
200	to 1,000 m3/h	0.24 %
60	to 200 m3/h	0.30 %

Calibration Date: November 17, 2008

Tested By: 
 Jeff Dahlin

Date of issue: November 17, 2008

Reviewed: 
 Wayne Haner

Certificate of Calibration

Certificate Number: 3051
Serial Number: 08-400217

Measurement Results

Qmax = 5583 m3/h
Qmin = 140 m3/h


Average Error -0.18
Adjusting N/A

Adjust Fact.(found) 1.0000
Adjust Fact.(left) see table

Uncorrected Results					Corrected Results			
MUT Flow (m3/hr)	Ref. Flow (m3/hr)	Velocity (m/s)	Deviation (%)	2xStDev (%)	Deviation (%)	Adj. Factor	Verification Pt. (%)	
5529.66	5549.81	30.31	-0.36	0.03	0.00	1.0036		
4169.89	4178.55	22.82	-0.21	0.03	0.00	1.0021		
2777.14	2780.58	15.19	-0.12	0.04	0.00	1.0012		
1401.61	1401.64	7.65	0.00	0.05	0.00	1.0000		
565.11	565.58	3.09	-0.08	0.06	0.00	1.0008		
274.32	274.70	1.50	-0.14	0.09	0.00	1.0014		
84.47	85.73	0.47	-1.47	0.22	0.00	1.0149		
172.18	171.51	0.94					0.39	Ver Pt1
3913.72	3914.08	21.38					-0.01	Ver Pt2

MUT is the abbreviation for Meter Under Test,
Deviations and Stdev are shown in % of reference reading
Deviation [%] = (Qmeter - Qreference)/Qreference*100

Adjustment Method: Piece-wise linearization



Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the issuing laboratory

TransCanada Calibrations Ltd.
Box 880
Ile des Chenes, Manitoba, Canada
R0A 0T0
(204) 878-4373

ISO/IEC 17025 Accredited
CLAS Certification No. 2005-04
SCC Accredited Lab No.: 591
Measurement Canada Accredited

Certificate of Calibration

Customer:	Lakeside Process Controls/EGD 5250 Orbitor Drive Mississauga, ON L4W 5G7 Gatineau	Certificate Number:	3457
		Work Order:	09-101-274-01
		Date of Calibration:	August 25, 2009
		Units:	Metric
Project:		End User MC Registration:	N/A
Meter Under Test			
Type:	Ultrasonic	Medium:	Natural Gas
Manufacturer:	Daniel	Pressure: (kPa)	6188.48
Model:	Senior Sonic	Temperature: (°C)	29.50
Serial Number:	09-260038	Density: (kg/m³)	45.17
Meter ID (m)	0.2545	Compressibility:	0.90385
Year of Manufacture:	2009	Direction:	Unidirectional
Meter Details			
CPU Board S/N:	9341	Pulses/m³:	3178.32
CPU Software Ver:	1.63	Meter output tested:	Frequency
Tag/Inspection Number:	N/A		
Pipe Spool S/N:	N/A		
Run Configuration:	UST 1670.05mm, CPA 50E, UST 2794mm, METER, DST 1670.05mm		
Flow Conditioning:	CPA 50E Type A		
Flow Conditioner S/N:	1466203		
Witness:	None		
Comments:	None		

Test Method: All calibrations are performed as outlined in TCC calibration procedure GC1-5

The actual flow of the gas meter is established with the aid of standard gas meters, pressure and temperature transmitters. In determination of this flow rate the pressure measurement point noted with Pr, is normatively established.

Results: The measurement results are stated on page 2 of 2. The type A uncertainty in the measurement is based on 2x the standard deviation (2XSTDEV). The total expanded uncertainty (U_{tot}) can be calculated as: $U_{tot} = \text{SQRT}[(U_{facility})^2 + (2XSTDEV)^2]$

Traceability: The Calibration Laboratory Assessment Service (CLAS) of the National Research Council of Canada (NRC) has assessed and certified specific calibration capabilities of this laboratory and traceability to the International System of Units (SI) or to standards acceptable to the CLAS program. The flow meter under test was calibrated in comparison with TransCanada Calibrations flow primary standards which are traceable to the SI and to recognized national metrology institutes. This certificate of calibration is issued in accordance with the conditions of certification granted by CLAS and the conditions of accreditation granted by the Standards Council of Canada (SCC). Neither CLAS nor SCC guarantee the accuracy of individual calibrations by accredited laboratories. TransCanada Calibrations is a Measurement Canada accredited meter verifier.

Facility Expanded Uncertainty, k=2:	Flow Rate	U _{facility} [% of reading]
1,000 to	55,000 m ³ /h	0.20 %
200 to	1,000 m ³ /h	0.24 %
60 to	200 m ³ /h	0.30 %

Calibration Date: August 25, 2009

Tested By: 
Jeff Dahlin

Date of Issue: August 25, 2009

Reviewed: 
Wayne Harner

Certificate of Calibration

Certificate Number: 3457
 Serial Number: 09-260038

Measurement Results

Qmax = 5583 m3/h
 Qmin = 140 m3/h


Average Error Adjusting: -0.55
N/A

Adjust Fact.(found): 1.0000
 Adjust Fact.(left): see table

Uncorrected Results					Corrected Results			
MUT Flow (m3/hr)	Ref. Flow (m3/hr)	Velocity (m/s)	Deviation (%)	2xStDev (%)	Deviation (%)	Adj. Factor	Verification Pt. (%)	
5468.43	5501.47	30.04	-0.60	0.03	0.00	1.0060		
4114.30	4136.70	22.59	-0.54	0.03	0.00	1.0054		
2732.97	2747.30	15.00	-0.52	0.04	0.00	1.0052		
1377.00	1382.48	7.55	-0.40	0.05	0.00	1.0040		
549.45	552.82	3.02	-0.61	0.05	0.00	1.0061		
364.79	367.61	2.01	-0.77	0.06	0.00	1.0077		
188.14	190.18	1.04	-1.07	0.10	0.00	1.0109		
276.22	276.18	1.51					0.01	Ver Pt1
914.57	914.52	4.99					0.00	Ver Pt2

MUT is the abbreviation for Meter Under Test,
 Deviations and Stdev are shown in % of reference reading
 $Deviation [\%] = (Q_{meter} - Q_{reference}) / Q_{reference} * 100$

Adjustment Method: Piece-wise linearization



Copyright of this certificate is owned by the issuing laboratory and may not be reproduced other than in full except with the prior written approval of the issuing laboratory.